



## Specification

### Human Serum Normal Pool

**Catalog Number:** HD1001  
**Volume:** 100mL, sterile-filtered  
**Storage:** ≤ -20°C  
**Retest Period:** Five years from the date of manufacture

Test	Specification
Physical Appearance	Clear, straw yellow solution
Osmolality	263 - 323mOsm/kg
Hemoglobin	< 26mg/dL
pH	7.0 – 8.0
Electrophoresis	Representative of normal human serum
Sterility (USP)	Sterile
Mycoplasma	Negative
Endotoxin	< 10EU/mL
Glucose	70 – 118mg/dL
Sodium	135 – 148mEq/L
Potassium	3.5 - 5.5mEq/L
Chloride	95 – 110mEq/L
Calcium	7 – 11.5mg/dL
Phosphorus	2.2 – 7.0mg/dL
Total Protein	6.0 – 8.0g/dL
Albumin	3.0 - 5.5g/dL
Total Bilirubin	Report
Cholesterol	80 – 200mg/dL
Triglyceride	30 – 160mg/dL
The serum is filtered through a sterile 0.1 micron cartridge.	
Each donor unit used in the preparation of this material was tested and found negative for HBsAG, HIV-1/2, anti-HCV, HCV-NAT, HIV 1-NAT, and a test for syphilis by FDA approved methods. Male donors blood type AB, collected in FDA licensed commercial donor establishments located in the US and processed in an FDA registered facility located in the US, in accordance with ethical standards.	

Approved:

Date: April 6, 2018

John Grubb, PhD  
Gojira Fine Chemicals  
Quality Manager  
Tel: 440-252-5397, Fax: 888-211-5523  
[docsupport@gojiraafc.com](mailto:docsupport@gojiraafc.com)