



## Specification

### Brucella Broth

**CAS Number:** N/A  
**Molecular Formula:** N/A  
**Molecular Weight:** N/A  
**Storage:** 15-30°C

Test	Specification
DCM Appearance:	homogeneous free flowing, light beige
PPM Appearance:	Clear, no to trace ppt, light yellow
pH (2.8 % Solution @25 oC):	6.8 - 7.2
Microorganism (Brucella Ovis ATCC 25840):	Poor to Good
Microorganism (Campylobacter Jejuni ATCC 29428):	Microorganism (Campylobacter Jejuni ATCC 29428): Poor to Good
Microorganism (Escherichia Coli ATCC 25922):	Microorganism (Escherichia Coli ATCC 25922): Very Good to Excellent
Microorganism (Neisseria meningitides ATCC 13090):	Microorganism (Neisseria meningitides ATCC 13090): Fair to very Good
Microorganism (Staphylococcus Aureus ATCC 25923):	Microorganism (Staphylococcus Aureus ATCC 25923): Very Good to Excellent
Microorganism (Streptococcus pneumoniae ATCC 6305):	Microorganism (Streptococcus pneumoniae ATCC 6305): Very Good to Excellent
Microorganism (Streptococcus pyogenes ATCC 19615):	Microorganism (Streptococcus pyogenes ATCC 19615): Very Good to Excellent

Approved:

*John Grubb*

Date: May 19, 2021



# Gojira Fine Chemicals, LLC

Quality Products & Services Meeting Customer Needs  
5386 Majestic Parkway, Suite 7, Bedford Heights, OH 44146  
[www.gojirafc.com](http://www.gojirafc.com) • [sales@gojirafc.com](mailto:sales@gojirafc.com)

---

John Grubb, PhD  
Gojira Fine Chemicals  
Quality Manager  
Tel: 440-252-5397, Fax: 888-211-5523  
[docsupport@gojirafc.com](mailto:docsupport@gojirafc.com)