



## Specification

### Columbia CNA Agar

**CAS Number:** N/A  
**Molecular Formula:** N/A  
**Molecular Weight:** N/A  
**Storage:** 15-30°C

Test	Specification
DCM Appearance:	Homogeneous, free flowing beige/red powder
PPM Appearance:	Light Beige/Red
pH @4.3% solution @ 25oC:	7.1 - 7.5
Microorganism -Pseudomonas Aeruginosa ATCC 27853:	Inhibited
Microorganisms (Proteus Mirabilis ATCC 12453):	Inhibited
Microorganism (Staphylococcus Aureus ATCC 25923):	Good Growth- Beta Hemolysis
Microorganism (Streptococcus pneumoniae ATCC 6305):	Good Growth- Alpha Hemolysis
Microorganism (Streptococcus pyogenes ATCC 19615):	Good Growth- Beta Hemolysis

Approved:

*John Grubb*

Date: May 19, 2021



# Gojira Fine Chemicals, LLC

Quality Products & Services Meeting Customer Needs  
5386 Majestic Parkway, Suite 7, Bedford Heights, OH 44146  
[www.gojirafc.com](http://www.gojirafc.com) • [sales@gojirafc.com](mailto:sales@gojirafc.com)

---

John Grubb, PhD  
Gojira Fine Chemicals  
Quality Manager  
Tel: 440-252-5397, Fax: 888-211-5523  
[docsupport@gojirafc.com](mailto:docsupport@gojirafc.com)